



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Jan 27, 2010 INITIALS: WJL OWNER: WASHINGTON STATE PATROL
DISTRICT: 3 ADDRESS: _____
LOCATION: FM SHOP COUNTY: _____
MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII
MODEL: 101 ID. NO: 24129 CAPACITY: 20000 lbs. SECTIONS: _____
MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19900	-100				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Certificate of Accuracy", NO. 12345-W-02.

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Wan La Tremere Date: Feb. 11, 2010

State of Washington
County of Shuster
Signed and sworn to (or affirmed before
me on Feb 11, 2010 by

Brenda A Biscay
(Signature of notary public)
Brenda A. Biscay
Printed Name of Notary
My appointment expires 3.12.12

